



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 12-Mar-2014

Lot # AG406901

Exp. Date

10-Mar-2016

Cyl. Type

108

Component

Ethanol
Nitrogen

Certified Concentration

82 ± 2 mg/100ml (223 ppm)
Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.

EB0010581
EB0010570
EB0010285
EB0010561
EB0010681

Concentration

391.8 ppm
259.8 ppm
209.0 ppm
103.7 ppm
52.22 ppm

Serial No.

EB0010603
EB0010559
EB0010595
EB0010562
EB0010579

Concentration

392.5 ppm
258.9 ppm
208.9 ppm
104.9 ppm
52.94 ppm

Analytical Method:

NDIR

Analyst: 
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01