



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Exclusive Supplier  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 12-Jul-2016

**Lot #** AG618701 **Model** 108cacd

**Exp. Date**

5-Jul-2018

**Cyl. Type**

108

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

82 ± 2 mg/100ml (223 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

**Analytical Method:**

NDIR

Analyst: \_\_\_\_\_

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**