



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Exclusive Supplier  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 22-Jun-2017

**Lot #** AG717101 **Model** 108cacd

**Exp. Date**

20-Jun-2019

**Cyl. Type**

108

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

82 ± 2 mg/100ml (223 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

| <u>Serial No.</u> | <u>Concentration</u> | <u>Serial No.</u> | <u>Concentration</u> |
|-------------------|----------------------|-------------------|----------------------|
| EB0010581         | 392.1 ppm            | EB0010603         | 393.0 ppm            |
| EB0010570         | 259.8 ppm            | EB0010559         | 258.2 ppm            |
| EB0010285         | 208.0 ppm            | EB0010595         | 208.3 ppm            |
| EB0010561         | 103.6 ppm            | EB0010562         | 104.2 ppm            |
| EB0010681         | 52.12 ppm            | EB0010579         | 52.81 ppm            |

**Analytical Method:**

NDIR

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01