Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 19-Jun-2019

Lot # AG916801 **Model** 108cacd

Exp. Date	Cyl. Type	Component	Certified Concentration
17-Jun-2021	108	Ethanol	82 ± 2 mg/100ml (223 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	**392.1 ppm**	**EB0010603**	**393.0 ppm**
EB0010570	**259.8 ppm**	**EB0010559**	**258.2 ppm**
EB0010285	**208.0 ppm**	**EB0010595**	**208.3 ppm**
EB0010561	**103.6 ppm**	**EB0010562**	**104.2 ppm**
EB0010681	**52.12 ppm**	**EB0010579**	**52.81 ppm**

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	**800.0 ppm**	**0056649**	**390.1 ppm**
CC234503	**253.0 ppm**	**0056662**	**150.2 ppm**

Analytical Method: **NDIR**

Digitally signed by Quality Control
Date: 2019.06.19 10:45:48 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release: Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07