Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 20-Mar-2025

Lot # AG507702 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
18-Mar-2027	108	Ethanol Nitrogen	82 ± 2 mg/100ml (223 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

Analytical Method: **NDIR**

Digitally signed by:Quality Control
Reason:Dry gas standard certification of analysis
Location:Airgas USA LLC (Lab)
Date:03.21.2025 06:57

Approved for Release: _____
Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07